

Publication Date: 09/26/2019 Last Updated: 09/26/2019

Abbott is monitoring developments related to recently published advisory (ICSA-19-211-01) identifying 11 IPNet vulnerabilities in Wind River’s VxWorks and other widely used Real Time Operating Systems (RTOSs). These vulnerabilities were reported by security researchers at Armis and are sometimes referred to as “Urgent/11”. RTOSs are used in a wide variety of products, including printers, routers, medical devices, firewalls, VOIP phones and industrial controllers. The vulnerabilities identified could allow unauthorized attackers to execute code remotely and take control of the system.

Most Abbott products are unaffected by this advisory. For those products that do use impacted firewalls or software, Abbott’s product security and quality teams are evaluating the vulnerabilities for potential impact including further actions and updates that may be required. There are currently no known exploits of Abbott products related to Urgent/11.

Affected Products

Abbott is providing the list below to assist customers in identifying Abbott products potentially impacted by Urgent/11. This list below is subject to change based on additional information related to impacted operating systems, operating system vendor actions and additional impacted products that are identified.

Product	Remediation
Third-party SonicWall Firewalls provided with the following Abbott products: ACCELERATOR APS ACCELERATOR a3600 ACCELERATOR p540 Alinity h Alinity ci-series Alinity s Alinity m ARCHITECT CELL-DYN Ruby CELL-DYN Sapphire m2000	Abbott's configuration of SonicWall Firewalls includes disabling remote management access from untrusted internet sources, which mitigates the vulnerabilities per SonicWall security advisory SNWLID-2019-0009 published 19 July 2019.
i-STAT Alinity	Abbott is developing a software update which is expected to be available to customers in November 2019.

Additional Information

Customers interested in additional information regarding patches, procedures or configuration changes on any Abbott products should contact their Abbott assigned account or customer support representative.

Abbott is committed to ensuring the safety and security our products. For more information on Abbott's product cybersecurity program, please go to: <https://www.abbott.com/policies/cybersecurity/our-commitment-to-cybersecurity.html>